

Ürün Kodu	Test	Ölçüm Aralığı	Analiz Sayısı	Raf Ömrü(Yıl)	Kit Tipi	Spectrophotometer	500 D	PF-12 <sup>Plus</sup>	PF-3 COD	PF-3 Drinking Water	PF-3 Fish	PF-3 Pool	PF-3 Soil
985098	Aluminum 07	0.02 - 0.70 mg/L Al3+	19	1	Tube	✓	✓	✓	-	-	-	-	-
985004	Ammonium 10	0,2 - 8,0 mg/L NH4-N 0,2 - 10,0 mg/L NH4+	20	1	Tube		✓	✓	-	-	-	-	✓
985008	Ammonium 100	4 - 80 mg/L NH4-N 5 - 100 mg/L NH4+	20	1	Tube	✓	✓	✓	-	-	-	-	-
985006	Ammonium 200	30 - 160 mg/L NH4-N 40 - 200 mg/L NH4+	20	1	Tube	✓	✓	✓	-	-	-	-	-
985002	Ammonium 2000	300 - 1600 mg/L NH4-N 400 - 2000 mg/L NH4+	20	1	Tube	✓	✓	✓	-	-	-	-	-
985003	Ammonium 3	0,04 - 2,30 mg/L NH4-N 0,05 - 3,00 mg/L NH4+	20	1	Tube	✓	✓	✓	-	-	✓	-	✓
985005	Ammonium 50	1 - 40 mg/L NH4-N 1 - 50 mg/L NH4+	20	1	Tube	✓	✓	✓	-	-	-	-	✓
985032	Anionic surfactants 4	0,20 - 4,00 mg/l MBAS	20	2	Tube	✓	✓	✓	-	-	-	-	-
985007	AOX 3	0.1 - 3.0 mg/L AOX 0.01 - 0.30 mg/L AOX	20	1	Tube	✓	✓	✓	-	-	-	-	-
985822	BOD <sub>5</sub>	2 - 3000 mg/L O <sub>2</sub>	25-50	2	Tube	✓	✓	✓	-	-	-	-	-
985825	BOD <sub>5</sub> -TT	0.5 - 3000 mg/L O <sub>2</sub>	22	2	Tube	✓	✓	✓	-	-	-	-	-
985014	Cadmium 2	0.05 - 2.00 mg/L Cd2+	10-19	1	Tube	✓	✓	✓	-	-	-	-	-
985015	Carbonate hardness 15	1.0 - 15.0 °d 0.4 - 5.4 mmol/L H+	20	1	Tube	✓	✓	✓	-	-	-	-	-
985034	Cationic surfactants 4	0,20 - 4,00 mg/l CTAB			Tube	✓	✓	✓	-	-	-	-	-
985019	Chloride 200	5 - 200 mg/L Cl-	20	1	Tube	✓	✓	✓	-	-	-	-	-
985021	Chloride 50	0,5 - 50 mg/L Cl-	20	1	Tube	✓	✓	✓	-	-	-	-	-
985017	Cholorine / Ozone 2	0.05 - 2.5 mg/L Cl <sub>2</sub> 0.05 - 2.0 mg/L O <sub>3</sub>	20	1	Tube	✓	✓	✓	-	✓	-	✓	-
985024	Chromate 5	0.1 - 4.0 mg/L CrO <sub>4</sub> <sup>2-</sup> 0.05 - 2.00 mg/L Cr(VI)  when using semi-micro cuvettes 50 mm (REF 919 50): 0.01 - 1.00 mg/L CrO <sub>4</sub> <sup>2-</sup> 0.005 - 0.500 mg/L Cr(VI)	20	2	Tube	✓	✓	✓	-	-	-	-	-

Ürün Kodu	Test	Ölçüm Aralığı	Analiz Sayısı	Raf Ömrü(Yıl)	Kit Tipi	Spectrophotometer	500 D	PF-12 <sup>Plus</sup>	PF-3 COD	PF-3 Drinking Water	PF-3 Fish	PF-3 Pool	PF-3 Soil
985018	Chlorine dioxide 5	0.15 - 5.00 mg/L ClO <sub>2</sub>	20	1	Tube	✓	✓	✓	-	-	-	-	-
985023	COD 10000	1.00 - 10.00 g/L O <sub>2</sub>	20	1	Tube	✓	✓	✓	✓	-	-	-	-
985029	COD 1500 (ISO 15705)	100 - 1500 mg/L O <sub>2</sub>	20	1	Tube	✓	✓	✓	✓	-	-	-	-
963029	COD 1500 Hg-free	100 - 1500 mg/L O <sub>2</sub>	20	1	Tube	✓	✓	✓	✓	-	-	-	-
985028	COD 15000 (ISO 15705)	1.0 - 15.0 g/L O <sub>2</sub>	20	1	Tube	✓	✓	✓	✓	-	-	-	-
985026	COD 160 (ISO 15705)	15 - 160 mg/L O <sub>2</sub>	20	1	Tube	✓	✓	✓	✓	-	-	-	-
963026	COD 160 Hg-free	15 - 160 mg/L O <sub>2</sub>	20	1	Tube	✓	✓	✓	-	-	-	-	-
985033	COD 300	50 - 300 mg/L O <sub>2</sub>	20	1	Tube	✓	✓	✓	-	-	-	-	-
985027	COD 40	2 - 40 mg/L O <sub>2</sub>	20	1	Tube	✓	✓	✓	✓	-	-	-	-
985011	COD 4000	400 - 4000 mg/L O <sub>2</sub>	20	1	Tube	✓	✓	✓	✓	-	-	-	-
985022	COD 60 (ISO 15705)	5 - 60 mg/L O <sub>2</sub>	20	1	Tube	✓	✓	✓	✓	-	-	-	-
985030	COD 600 (ISO 15705)	50 - 600 mg/L O <sub>2</sub>	20	1	Tube	✓	✓	✓	✓	-	-	-	-
985012	COD 60000	5.0 - 60.0 g/L O <sub>2</sub>	20	1	Tube	✓	✓	✓	✓	-	-	-	-
985038	COD HR 1500 (ISO 15705)	20 - 1500 mg/L O <sub>2</sub>	20	1	Tube	✓	✓	✓	✓	-	-	-	-
985036	COD LR 150 (ISO 15705)	3 - 150 mg/L O <sub>2</sub>	20	1	Tube	✓	✓	✓	✓	-	-	-	-
985053	Copper 5	0.10 - 7.00 mg/L Cu <sup>2+</sup>	20	2	Tube	✓	✓	✓	-	-	-	-	-
985054	Copper 7	0.10 - 7.00 mg/L Cu <sup>2+</sup>	20	2	Tube	✓	✓	✓	-	-	-	-	-
985031	Cyanide 08	0.02 - 0.80 mg/L CN- when using semi-micro cuvettes 50 mm (REF 919 50): 0.005 - 0.100 mg/L CN-	20	1	Tube	✓	✓	✓	-	-	-	-	-
985035	DEHA 1 (Diethylhydroxylamine)	0.05 - 1.00 mg/L DEHA	20	1	Tube	✓	✓	✓	-	-	-	-	-
985838	Ethanol 1000	0.10 - 1.00 g/L EtOH 0.013 - 0.13 Vol.% EtOH	23	2	Tube	✓	✓	✓	-	-	-	-	-
985040	Fluoride 2	0.1 - 2.0 mg/L F-	20	1,5	Tube	✓	✓	✓	-	-	-	-	-

Ürün Kodu	Test	Ölçüm Aralığı	Analiz Sayısı	Raf Ömrü(Yıl)	Kit Tipi	Spectrophotometer	500 D	PF-12 <sup>Plus</sup>	PF-3 COD	PF-3 Drinking Water	PF-3 Fish	PF-3 Pool	PF-3 Soil
985046	Formaldehyde 10	0.20 - 10.00 mg/L HCHO 0.02 - 1.00 mg/L HCHO	20	2	Tube	✓	✓	✓	-	-	-	-	-
985041	Formaldehyde 8	0.1 - 8.0 mg/L HCHO	20	2	Tube	✓	✓	✓	-	-	-	-	-
985043	Hardness 20	1.25 - 25.00 °e 5 - 50 mg/L Mg <sup>2+</sup> 10 - 100 mg/L Ca <sup>2+</sup>	20	1,5	Tube	✓	✓	✓	-	-	-	-	-
985044	Hardness Ca/Mg	1.25 - 25.00 °e 5 - 50 mg/L Mg <sup>2+</sup> 10 - 100 mg/L Ca <sup>2+</sup>	20	1,5	Tube	✓	-	✓	-	-	-	-	-
985057	HC 300 (Hydrocarbons)	0.5 - 5.6 mg/l HC 30 - 300 mg/kg HC	20	1	Tube	✓	✓	✓	-	-	-	-	-
985037	Iron 3	0.10 - 3.00 mg/L Fe  when using semi-micro cuvettes 50 mm (REF 919 50): 0.02 - 1.00 mg/L Fe	20	1	Tube	✓	✓	✓	-	-	-	-	-
985009	Lead 5	0.10 - 5.00 mg/L Pb <sup>2+</sup>	20	1	Tube	✓	✓	✓	-	-	-	-	-
985058	Manganese 10	0.1 - 10.0 mg/L Mn when using semi-micro cuvettes 50 mm (REF 91950): 0.02 - 2.00 mg/L Mn	20	1,5	Tube	✓	✓	✓	-	-	-	-	-
985859	Methanol 15	0.2 - 15.0 mg/L MeOH	23	1	Tube	✓	✓	✓	-	-	-	-	-
985056	Molybdenum 40	1.0 - 40.0 mg/l Mo(VI) 1.6 - 65.0 mg/l MoO <sub>4</sub> <sup>2-</sup>	20	2	Tube	✓	✓	✓	-	-	-	-	-
985071	Nickel 4	0,10 - 7,00 mg/L Ni <sup>2+</sup>	20	2	Tube	✓	✓	✓	-	-	-	-	-
985061	Nickel 7	0,10 - 7,00 mg/L Ni <sup>2+</sup>	20	2	Tube	✓	-	-	-	-	-	-	-
985066	Nitrate 250	4 - 60 mg/L NO <sub>3</sub> -N 20 - 250 mg/L NO <sub>3</sub> -	20	2	Tube	✓	✓	✓	-	-	-	-	✓
985064	Nitrate 50	0.3 - 22.0 mg/L NO <sub>3</sub> -N 2 - 100 mg/L NO <sub>3</sub> -	20	2	Tube	✓	✓	✓	-	-	-	-	-
985065	Nitrate 8	0.30 - 8.00 mg/L NO <sub>3</sub> -N 1.3 - 35.0 mg/L NO <sub>3</sub> -	20	2	Tube	✓	✓	✓	-	-	-	-	-
985068	Nitrite 2	0.003 - 0.460 mg/L NO <sub>2</sub> -N 0.02 - 1.50 mg/L NO <sub>2</sub> -	20	1	Tube	✓	✓	✓	-	-	-	-	-
985069	Nitrite 4	0.1 - 4.0 mg/L NO <sub>2</sub> -N 0.3 - 13.0 mg/L NO <sub>2</sub> -	20	1,5	Tube	✓	✓	✓	-	-	-	-	-

Ürün Kodu	Test	Ölçüm Aralığı	Analiz Sayısı	Raf Ömrü(Yıl)	Kit Tipi	Spectrophotometer	500 D	PF-12 <sup>Plus</sup>	PF-3 COD	PF-3 Drinking Water	PF-3 Fish	PF-3 Pool	PF-3 Soil
985047	Nonionic surfactants 15	0,3 - 15,0 mg/l Triton® X-100	20	2	Tube	✓	✓	✓	-	-	-	-	-
985052	org. Complexing agents 10	0.5 - 10.0 mg/L IBIK	10-19	1	Tube	✓	✓	✓	-	-	-	-	-
985050	Organic acids 3000	30 - 3000 mg/L CH <sub>3</sub> COOH 0.5 - 50.0 mmol/L CH <sub>3</sub> COOH	20	1,5	Tube	✓	✓	✓	-	-	-	-	-
985076	ortho- and total Phosphate 1	0.01 - 1.50 mg/L P (PO <sub>4</sub> -P) 0.03 - 5.0 mg/L PO <sub>4</sub> 3-	20	1	Tube	✓	✓	✓	-	-	-	-	-
985080	ortho- and total Phosphate 15	0.30 - 15.00 mg/L P (PO <sub>4</sub> -P) 1.0 - 45.0 mg/L PO <sub>4</sub> 3-	20	1	Tube	✓	✓	✓	-	-	-	-	✓
985055	ortho- and total Phosphate 45	5.0 - 50.0 mg/L P (PO <sub>4</sub> -P) 15 - 150 mg/L PO <sub>4</sub> 3-	20	1	Tube	✓	✓	✓	-	-	-	-	-
985081	ortho- and total Phosphate 5	0.20 - 5.00 mg/L P (PO <sub>4</sub> -P) 0.5 - 15.0 mg/L PO <sub>4</sub> 3-	20	1	Tube	✓	✓	✓	-	-	-	-	✓
985079	ortho- and total Phosphate 50	10.0 - 50.0 mg/L P (PO <sub>4</sub> -P) 30 - 150 mg/L PO <sub>4</sub> 3-	19	3	Tube	✓	✓	✓	-	-	-	-	-
985095	ortho- and total Phosphate LR 1	0.05 - 0.50 mg/L P (PO <sub>4</sub> -P) 0.2 - 1.5 mg/L PO <sub>4</sub> 3-	20	1	Tube	✓	✓	✓	-	-	-	-	-
985082	Oxygen 12	0.5 - 12.0 mg/l O <sub>2</sub>	22	2	Tube	✓	✓	✓	-	-	-	-	-
985871	Peroxide 2	0.03 - 2.00 mg/l H <sub>2</sub> O <sub>2</sub>	10-19	1	Tube	✓	✓	✓	-	-	-	-	-
91872	pH 6.5-8.2	pH 6.5 - 8.2	100	1,5	Tube	✓	✓	✓	-	✓	-	✓	-
985074	Phenolic Index 5	0.2 - 5.0 mg/l Phenolic-Index	20	1,5	Tube	✓	✓	✓	-	-	-	-	-
985070	POC 200	20 - 200 mg/l POC AS 2020 20 - 200 mg/l POC HS 2020 20 - 200 mg/l Polystabil® DK 2 - 40 mg/l Polystabil® KWI	20	1,5	Tube	✓	✓	✓	-	-	-	-	-
985045	Potassium 50	2 - 50 mg/L K <sup>+</sup>	20	2	Tube	✓	✓	✓	-	-	-	-	✓
985084	Residual hardness 1	0.02 - 1.00 °d 0.004 - 0.180 mmol/l	20	1	Tube	✓	✓	✓	-	-	-	-	-
985049	Silver 3	0.20 - 3.00 mg/l Ag	20	1,5	Tube	✓	✓	✓	-	-	-	-	-
985085	Starch 100	5 - 100 mg/l starch	19	1	Tube	✓	✓	✓	-	-	-	-	-
985087	Sulfate 1000	200 - 1000 mg/L SO <sub>4</sub> 2-	20	3	Tube	✓	✓	✓	-	-	-	-	-

Ürün Kodu	Test	Ölçüm Aralığı	Analiz Sayısı	Raf Ömrü(Yıl)	Kit Tipi	Spectrophotometer	500 D	PF-12 <sup>Plus</sup>	PF-3 COD	PF-3 Drinking Water	PF-3 Fish	PF-3 Pool	PF-3 Soil
985062	Sulfate LR 200	20 - 200 mg/L SO42-	20	3	Tube	✓	✓	✓	-	-	-	-	-
985073	Sulfide 3	0.05 - 3.00 mg/L S2-	20	3	Tube	✓	✓	✓	-	-	-	-	-
985089	Sulfite 10	0,05 - 10,0 mg/l SO32-	20	1	Tube	✓	✓	✓	-	-	-	-	-
985090	Sulfite 100	5 - 100 mg/l SO32-	19	1	Tube	✓	✓	✓	-	-	-	-	-
985091	Thiocyanide 50	0.5 - 50.0 mg/l SCN-	20	2	Tube	✓	✓	✓	-	-	-	-	-
985097	Tin 3	0.10 - 3.00 mg/l Sn	18	1	Tube	✓	✓	✓	-	-	-	-	-
985075	TOC 30	2.0 - 30.0 mg/L C	20	1	Tube	✓	✓	✓	-	-	-	-	-
985078	TOC 300	20 - 300 mg/L C	20	1	Tube	✓	✓	✓	-	-	-	-	-
985059	total Chromium 2	0.05 - 2.00 mg/L Cr when using semi-micro cuvettes 50 mm (REF 919 50): 0.005 - 0.500 mg/L Cr	20	2	Tube	✓	✓	✓	-	-	-	-	-
985083	total Nitrogen TN <sub>b</sub> 22	0.5 - 22.0 mg/L N	20	1	Tube	✓	✓	✓	-	-	-	-	-
985088	total Nitrogen TN <sub>b</sub> 220	5 - 220 mg/l N	20	1	Tube	✓	✓	✓	-	-	-	-	-
985092	total Nitrogen TN <sub>b</sub> 60	3 - 60 mg/L N	20	1	Tube	✓	✓	✓	-	-	-	-	-
985890	TTC/Sludge activity	5 - 150 µg TPF 0.050 - 2.300 E	20	2	Tube	✓	✓	✓	-	-	-	-	-
Test9-06	Turbidity	0.1-1000 NTU	-	-	Tube	✓	-	✓	-	-	-	-	-
985096	Zinc 4	0.10 - 4.00 mg/l Zn2+	20	1	Tube	✓	✓	✓	-	-	-	-	-
985001	Zirconium 100	5 - 100 mg/L Zr	20	3	Tube	✓	✓	✓	-	-	-	-	-

Ürün Kodu	Test	Ölçüm Aralığı	Analiz Sayısı	Raf Ömrü(Yıl)	Kit Tipi	Spectrophotometer	500 D	PF-12 <sup>Plus</sup>	PF-3 COD	PF-3 Drinking Water	PF-3 Fish	PF-3 Pool	PF-3 Soil
91802	Aluminum	0.01 - 1.00 mg/L Al <sup>3+</sup>	250	2	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91805	Ammonium	0.01 - 2.0 mg/L NH <sub>4</sub> -N 0.01 - 2.5 mg/L NH <sub>4</sub> <sup>+</sup>	100	1	Standart/Cuvette	✓	✓	-	-	-	-	-	-
918131	Cadmium	0.002 - 0.50 mg/L Cd <sup>2+</sup>	25	1,5	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91820	Chloride	0.2 - 125 mg/L Cl <sup>-</sup>	250	1	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91816	Chlorine	0.02 - 10.0 mg/L Cl <sub>2</sub>	250	3	Standart/Cuvette	✓	✓	-	-	-	-	-	-
918163	Chlorine dioxide	0.04 - 4.00 mg/L ClO <sub>2</sub>	50	1,5	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91825	Chromate	10 mm cuvette: 0.1 - 6.0 mg/l CrO <sub>4</sub> <sup>2-</sup> / 0.1 - 3.0 mg/l Cr(VI) 50 mm cuvette: 0.01 - 1.00 mg/l CrO <sub>4</sub> <sup>2-</sup> / 0.01 - 0.45 mg/l Cr(VI)	250	2	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91851	Cobalt	0.002 - 0.70 mg/L Co <sup>2+</sup>	250	2	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91853	Copper	10 mm cuvette: 0.1 - 10.0 mg/l Cu <sup>2+</sup> 50 mm cuvette: 0.01 - 2.00 mg/l Cu <sup>2+</sup>	250	2	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91830	Cyanide	0.001 - 0.50 mg/L CN <sup>-</sup>	250	1	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91832	Detergents, anionic	10 mm cuvette: 0.1 - 5.0 mg/l MBAS 50 mm cuvette: 0.02 - 1.00 mg/l MBAS	40	3	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91834	Detergents, cationic	10 mm cuvette: 0.2 - 5.0 mg/l CTAB 50 mm cuvette: 0.05 - 3.00 mg/l CTAB	40	3	Standart/Cuvette	✓	✓	-	-	-	-	-	-
918142	Fluoride	0.05 - 2.00 mg/L F <sup>-</sup>	500	2	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91844	Hydrazine	0.002 - 1.50 mg/L N <sub>2</sub> H <sub>4</sub>	250	1	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91836	Iron	10 mm cuvette: 0.1 - 15.0 mg/l Fe 50 mm cuvette: 0.01 - 2.00 mg/l Fe	250	3	Standart/Cuvette	✓	✓	-	-	-	-	-	-
918101	Lead	0.005 - 1.00 mg/L Pb <sup>2+</sup>	50	1,5	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91860	Manganese	10 mm cuvette: 0.1 - 10.0 mg/l Mn 50 mm cuvette: 0.01 - 2.00 mg/l Mn	250	3	Standart/Cuvette	✓	✓	-	-	-	-	-	-

Ürün Kodu	Test	Ölçüm Aralığı	Analiz Sayısı	Raf Ömrü(Yıl)	Kit Tipi	Spectrophotometer	500 D	PF-12 <sup>Plus</sup>	PF-3 COD	PF-3 Drinking Water	PF-3 Fish	PF-3 Pool	PF-3 Soil
91862	Nickel	10 mm cuvette: 0.1 - 10.0 mg/l Ni2+ 50 mm cuvette: 0.01 - 2.00 mg/l Ni2+	250	2	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91865	Nitrate	0.1 - 30.0 mg/L NO3-N 0.5 - 140 mg/L NO3-	100	2	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91863	Nitrate Z	0.02 - 1.00 mg/L NO3-N 0.1 - 5.0 mg/L NO3-	440	1,5	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91867	Nitrite	10 mm cuvette: 0.01 - 0.30 mg/l NO2-N / 0.10 - 1.00 mg/l NO2- 50 mm cuvette: 0.002 - 0.100 NO2-N / 0.005 - 0.250 NO2-	220	1,5	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91885	Ozone	10 mm cuvette: 0.10 - 1.50 mg/L O3 50 mm cuvette: 0.01 - 1.00 mg/L O3 16 mm cuvette: 0.05 - 1.50 mg/L O3 (only PF-12 and PF-12Plus)	200	1	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91875	Phenol	0.01 - 7.0 mg/L Phenol	500	3	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91877	ortho-Phosphate	0.04 - 6.5 mg/l PO4-P 0.1 - 20.0 mg/l PO43-	500	3	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91878	ortho-Phosphate	0.2 - 17 mg/L PO4-P 0.5 - 50 mg/L PO43-	500	3	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91848	Silica	0.005 - 10.0 mg/L SiO2 0.002 - 5.0 mg/L Si	250	3	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91888	Sulfide	0.01 - 3.0 mg/L S2-	250	3	Standart/Cuvette	✓	✓	-	-	-	-	-	-
91895	Zinc	10 mm cuvette: 0.1 - 3.0 mg/L Zn2+ 50 mm cuvette: 0.02 - 1.50 mg/L Zn2+	250	3	Standart/Cuvette	✓	✓	-	-	-	-	-	-

Ürün Kodu	Test	Ölçüm Aralığı	Analiz Sayısı	Raf Ömrü(Yıl)	Kit Tipi	Spectrophotometer	500 D	PF-12 <sup>Plus</sup>	PF-3 COD	PF-3 Drinking Water	PF-3 Fish	PF-3 Pool	PF-3 Soil
92507	AOX 3	1.0 mg/L AOX	20	1	Single Ref. Standard								
92582	BOD5	210 mg/L O2	10	1	Single Ref. Standard								
92517	Chlorine	1.00 mg/L Cl2 0.80 mg/L Cl2	30	1	Single Ref. Standard								
92524	Chromate	2.0 mg/L CrO42- 0.90 mg/L Cr 0.40 mg/L CrO42-	30	1	Single Ref. Standard								
92522	COD 60	30 mg/L O2 30 mg/L O2	15	1	Single Ref. Standard								
92526	COD 160	100 mg/L O2 100 mg/L O2 100 mg/L O2	15	1	Single Ref. Standard								
92529	COD 1500	400 mg/L O2 400 mg/L O2 400 mg/L O2	15-30	1	Single Ref. Standard								
92528	COD 15000	4.00 g/L O2 4.0 g/L O2	30-150	1	Single Ref. Standard								
92568	Nitrite	0.060 mg/L NO2-N 0.30 mg/L NO2-N 2.1 mg/L NO2-N	15-150	1	Single Ref. Standard								
92576	ortho-Phosphate	1.00 mg/L PO4-P 0.25 mg/L PO4-P 0.20 mg/L PO4-P	15	1	Single Ref. Standard								
92562	Sulfate	120 mg/L SO42-	15	1	Single Ref. Standard								
92590	Sulfite	50 mg/L SO32-	15	1	Single Ref. Standard								
92575	TOC 30	10 mg/L C	15	1	Single Ref. Standard								
92578	TOC 300	100 mg/L C	15	1	Single Ref. Standard								



Ürün Kodu	Test	Ölçüm Aralığı	Analiz Sayısı	Raf Ömrü(Yıl)	Kit Tipi	Spectrophotometer	500 D	PF-12 <sup>Plus</sup>	PF-3 COD	PF-3 Drinking Water	PF-3 Fish	PF-3 Pool	PF-3 Soil
925011	Sewage outflow 1 Multistandard for the determination of COD, NO3-N, NH4-N, total N and total P	-	12-120	1	Multi Ref. Standard								
925010	Sewage outflow 2 Multistandard for determination of Ammonium, COD, Nitrate, total P and total N	-	12-120	1	Multi Ref. Standard								
925012	Sewage influx Multistandard for the determination of COD, NO3-N, NH4-N,	-	30-300	1	Multi Ref. Standard								
925013	Sewage water Multistandard for the determination of COD, NO3-N, NH4-N and ortho P and total P	-	15-300	1	Multi Ref. Standard								
925015	Metals 1 Multistandard for the determination of Cadmium, Chloride, Chromium, Iron, Fluoride, Sulfate and	-	15-60	1	Multi Ref. Standard								
925016	Metals 2 Multistandard for the determination of Lead, Potassium,	-	15	1	Multi Ref. Standard								

Ürün Kodu	Test	Ölçüm Aralığı	Analiz Sayısı	Raf Ömrü(Yıl)	Kit Tipi	Spectrophotometer	500 D	PF-12 <sup>Plus</sup>	PF-3 COD	PF-3 Drinking Water	PF-3 Fish	PF-3 Pool	PF-3 Soil
925018	Drinking water Multistandard for the determination of Aluminum, Ammonium, Chloride, Iron, Manganese and Sulphate	-	15-30	1	Multi Ref. Standard								

Ürün Kodu	Test	Ölçüm Aralığı	Analiz Sayısı	Raf Ömrü(Yıl)	Kit Tipi	Spectrophotometer	500 D	PF-12 <sup>Plus</sup>	PF-3 COD	PF-3 Drinking Water	PF-3 Fish	PF-3 Pool	PF-3 Soil
936220	Free chlorine	0.03 - 6.00 mg/L Cl <sub>2</sub>	100	5	Pillow	✓	-	✓	-	✓	✓	✓	-
936220.1	Free chlorine	0.03 - 6.00 mg/L Cl <sub>2</sub>	1000	5	Pillow	✓	-	✓	-	✓	✓	✓	-
936221	Total chlorine / ozone	0.03 – 6.00 mg/L Cl <sub>2</sub> 0.03 – 4.00 mg/L O <sub>3</sub>	100	5	Pillow	✓	-	✓	-	✓	-	✓	-
936221.1	Total chlorine / ozone	0.03 – 6.00 mg/L Cl <sub>2</sub> 0.03 – 4.00 mg/L O <sub>3</sub>	100	5	Pillow	✓	-	✓	-	✓	-	✓	-
936222	pH	pH 6.2 - 8.2	100	5	Pillow	✓	-	✓	-	✓	-	✓	-
936223	Sulfate	15 – 200 mg/L SO <sub>4</sub> <sup>2-</sup>	100	5	Pillow	✓	-	✓	-	✓	-	✓	-
936224	Silica LR	0.02-2.10 mg/L SiO <sub>2</sub> 0.01-1.0 mg/L Si	100	3	Pillow	✓	-	✓	-	✓	✓	✓	-
936225	Silica HR	2-210 mg/L SiO <sub>2</sub> 1-100 mg/L Si	100	3	Pillow	✓	-	✓	-	✓	✓	✓	-
931204	Alkalinity (total)	0.3–14 °d 0.10–5.00 mmol/L H <sup>+</sup> 0.4–17.5 °e 0.5–25.0 ° 5–250 mg/L CaCO <sub>3</sub>	100	1	Viscolor	-	-	✓	-	✓	-	✓	-
931008	Ammonium 3	0.1–2.0 mg/L NH <sub>4</sub> -N 0.1–2.5 mg/L NH <sub>4</sub> <sup>+</sup>	50	1,5	Viscolor	-	-	✓	-	-	✓	-	✓
931010	Ammonium 15	0.5–8.0 mg/L NH <sub>4</sub> <sup>+</sup>	50	1,5	Viscolor	-	-	✓	-	-	-	-	-
931211	Bromine	0.10-13.00 mg/L Br <sub>2</sub>	200	2	Viscolor	-	-	✓	-	✓	-	✓	-
931035	Chlorine 1, free + total	0.05 - 2.00 mg/L Cl <sub>2</sub>	150	2	Viscolor	-	-	✓	-	✓	-	✓	-
931015	Chlorine 2, free + total	0.10 - 2.00 mg/L Cl <sub>2</sub>	150	1,5	Viscolor	-	-	✓	-	✓	-	✓	-
931016	free Chlorine 2	0.10 - 2.00 mg/L Cl <sub>2</sub>	150	1,5	Viscolor	-	-	✓	-	✓	-	✓	-
931217	Chlorine 6, free + total	0.05 - 6.00 mg/L Cl <sub>2</sub>	200	2	Viscolor	-	-	✓	-	✓	✓	✓	-
931219	free Chlorine 6	0.05 - 6.00 mg/L Cl <sub>2</sub>	400	2	Viscolor	-	-	✓	-	✓	-	✓	-
931018	Chloride	1–50 mg/L Cl <sup>-</sup>	90	1	Viscolor	-	-	✓	-	-	-	-	-
931021	Chlorine dioxide	0.20 - 3.80 mg/L ClO <sub>2</sub>	150	1,5	Viscolor	-	-	✓	-	-	-	-	-
931020	Chromium(VI)	0.02 - 0.05 mg/L Cr(VI)	140	1,5	Viscolor	-	-	✓	-	-	-	-	-
931037	Copper	0.1 -5.0 mg/L Cu <sup>2+</sup>	100	2	Viscolor	-	-	✓	-	-	✓	-	-
931022	Cyanide	0.01 - 0.20 mg/L CN <sup>-</sup>	100	1	Viscolor	-	-	✓	-	-	-	-	-
931023	Cyanuric acid	10 - 100 mg/L Cya	100	1,5	Viscolor	-	-	✓	-	✓	-	✓	-
931227	Fluoride	0.1 - 2.0 mg/L F <sup>-</sup>	150	1,5	Viscolor	-	-	✓	-	-	-	✓	-
931030	Hydrazine	0.05 - 0.40 mg/L N <sub>2</sub> H <sub>4</sub>	130	1	Viscolor	-	-	✓	-	-	-	-	-
931025	Iron 1	0.04 - 2.00 mg/L Fe	200	2	Viscolor	-	-	✓	-	✓	✓	✓	-

Ürün Kodu	Test	Ölçüm Aralığı	Analiz Sayısı	Raf Ömrü(Yıl)	Kit Tipi	Spectrophotometer	500 D	PF-12 <sup>Plus</sup>	PF-3 COD	PF-3 Drinking Water	PF-3 Fish	PF-3 Pool	PF-3 Soil
931026	Iron 2	0.04 - 2.00 mg/L Fe	100	2	Viscolor	-	-	✓	-	✓	✓	✓	-
931038	Manganase	0.1 - 5.0 mg/L Mn	70	1,5	Viscolor	-	-	✓	-	-	-	-	-
931040	Nickel	0.04 - 5.00 mg/L Ni+2	150	1,5	Viscolor	-	-	✓	-	-	-	-	-
931041	Nitrate	4 - 60 mg/L NO3-	110	1,5	Viscolor	-	-	✓	-	-	✓	-	✓
931044	Nitrite	0.02 - 0.50 mg/L NO2-	120	1,5	Viscolor	-	-	✓	-	-	✓	-	-
931088	Oxygen	1-8 mg/L O2	50	1	Viscolor	-	-	✓	-	-	✓	-	-
931270	pH 6.0-8.2	pH 6.1-8.4	150	1,5	Viscolor	-	-	✓	-	✓	✓	✓	-
931084	Phosphate	0.2-5.0 mg/L PO4-P	80	3	Viscolor	-	-	✓	-	-	✓	-	✓
931032	Potassium	2-25 mg/L K+	60	3	Viscolor	-	-	✓	-	-	-	-	✓
931033	Silica	0.2-3.0 mg/L SiO2	80	3	Viscolor	-	-	✓	-	-	✓	-	-
931234	Silica HR 200	10-200 mg/L SiO2	100	3	Viscolor	-	-	✓	-	✓	-	✓	-
931092	Sulfate	20-200 mg/L SO4	100	3	Viscolor	-	-	✓	-	-	-	-	-
931094	Sulfide	0.05-0.80 mg/L S2-	90	3	Viscolor	-	-	✓	-	-	-	-	-
931098	Zinc	0.1-3.0 mg/L Zn+2	120	1	Viscolor	-	-	✓	-	-	-	-	-

[SU ANALİZ KİTLERİ VE CİHAZLARIMIZI HEMEN ŞİMDİ İNCELEYİN](#)