



SPECIFICATION SHEET

Product Name : Buffer Solution pH 2.00 ± 0.02 at 20°C
(Glycine, Hydrochloric Acid, Sodium Chloride)
Product Number : 908.B02
Brand : ISOLAB
CAS Number : -
Formula : -
Formula Weight : -

TEST NAME	SPECIFICATION	UNIT
pH	1.98 - 2.02 at 20°C	-
Measured with a combination glass electrode after multiple point calibration with reference materials. It is certified traceable to the following National Institute of Standards and Technology (USA), SRM 189c Potassium Tetraoxalate and SRM 185i Potassium Hydrogen Phthalate.		

Quality Management
Isolab Laborgerate GmbH

