



**SPECIFICATION SHEET**

**Product Name** : Buffer Solution pH 11.00 ± 0.05 at 20°C  
(Glycine, Sodium Chloride, Sodium Hydroxide)  
**Product Number** : 908.B11  
**Brand** : ISOLAB  
**CAS Number** : -  
**Formula** : -  
**Formula Weight** : -

TEST NAME	SPECIFICATION	UNIT
pH	10.95 - 11.05 at 20°C	-
Measured with a combination glass electrode after multiple point calibration with reference materials. It is certified traceable to the following National Institute of Standards and Technology (USA), SRM 191d-I Sodium Bicarbonate, SRM 191d-II Sodium Carbonate, SRM 186-I-g Potassium Dihydrogen Phosphate and SRM 186-II-g Disodium Hydrogen Phosphate.		

Quality Management  
Isolab Laborgerate GmbH

